



भौतिकी संस्थान

भुवनेश्वर

BY SPEED POST

Institute of Physics

Bhubaneswar - 751005, India

(परमाणु ऊर्जा विभाग, भारत सरकार का एक स्वायत्त अनुसंधान संस्थान)

(An Autonomous Research Institute of Department of Atomic Energy, Govt. of India)

QUOTATION CALL NOTICE

Ref.No. IOP/PUR/PRO/INSP/2021-22/ <u>594</u>	Rate quotation for:
Date: <u>08/07/2021</u>	Rotary Vacuum Pump

To:

M/s

Dear Sir,

Sealed quotation is invited for supply of Rotary Vacuum PUmp as per specification, terms and conditions noted below.

Sl.	Description/Specification	Qty	Price
1	Rotary Vacuum Pump for single zone furnace	01	
2	SS below with KF25 END FLANGE	01	
3	Molecular distilled Oil	01	

TERMS & CONDITIONS:

1. Intending supplier should be a registered dealer having TIN/GST Registration No. and should submit a copy of the same.
2. Quotation is to be submitted by **post/courier** only in sealed envelope **with superscription "Rate Quotation for supply of "Rotary Vacuum Pump" and addressed to the Registrar, Institute of physics, P.O: Sainik School, Bhubaneswar-751005.**
3. The quotation should be submitted within **25 days** from the date of issue of this letter in the office of the Registrar along with copies of the aforesaid certificates and authorization from OEM if any failing which the same will be rejected.
4. Rate quoted should remain valid for ninety days from the date of submission of the quotation.
5. Price quoted should be FOR destination at IOP site basis.
6. The items must be warranted against any functional defects.
7. Taxes and levies if applicable will be paid extra, which should be specified separately.
8. 100% payment will be made on delivery of goods in satisfactory conditions and against submission of bill/invoice in duplicate.

भवदीय/Yours faithfully


रजिस्टार/ REGISTRAR

डा.घ. सैनिक स्कूल, भुवनेश्वर - 751 005, भारत

P.O. : SAINIK SCHOOL, BHUBANESWAR - 751 005, INDIA

दूरभाष : 0674-2306444, 2306555, 2306666 फैक्स / Fax : 0674-2300142 वेब / Web : <http://www.iopb.res.in>

Pump specifications:

Single phase upto 220V, 50Hz

Ultimate vacuum: 2×10^{-3} mbar (w/o gas ballast)

: 5×10^{-2} mbar (with gas ballast)

Pumping speed: 12 m³/h

Water vapor pumping capacity: 0.29 kg/h

Inlet connection: KF 25

Outlet connection: 15 mm Dia nozzle