



भौतिकी संस्थान
भुवनेश्वर
Institute of Physics
Bhubaneswar - 751005, India

(परमाणु ऊर्जा विभाग, भारत सरकार का एक स्वायत्त अनुसंधान संस्थान)
(An autonomous research institution of Dept. of Atomic Energy, Govt. of India)

QUOTATION CALL NOTICE:

No.IOP/PUR/APEX(DIR)M&E/2022-23/ <u>127</u>	Rate quotation for:
Date: <u>02/02/2023</u>	FUME HOOD

To:

M/s

Dear Sir,


Sealed quotation is invited for Supply of Fume Hood as per specification, terms and conditions noted below.

Sl. No.	Description/Specification	Quantity	Price (Rs.)
1	FUME HOOD (Specification attached in a separate sheet)	01 No.	

TERMS & CONDITIONS:

1. Intending supplier should be a registered dealer having GST Registration No. and should submit a copy of the same.
2. Quotation is to be submitted by **post/courier** only in **sealed envelope with superscription "Fume Hood"** and addressed to the **Registrar, Institute of physics, P.O: Sainik School, Bhubneswar-751005.**
3. The quotation should be submitted within **21 days** from the date of issue of this letter in the office of the Registrar along with copies of the aforesaid certificates if any failing which the same will be rejected.
4. Rate quoted should remain valid for **90 days** from the date of submission of the quotation.
5. Price quoted should be FOR destination at IOP site.
6. Taxes and levies if applicable will be paid extra, which should be specified separately.
7. **No advance payment to be made.** 100% payment will be made on delivery, installation (if required) of goods in satisfactory conditions and against submission of bill/invoice in duplicate.

भवदीय/ Yours faithfully


3/1/23
रजिस्ट्रार/REGISTRAR

डा.घ.- सैनिक स्कूल, भुवनेश्वर - 751 005, भारत

P.O. : SAINIK SCHOOL, BHUBANESWAR-751005, INDIA

दूरभाष / Phone : 0674-2300637, 2301058, 2301823, फैक्स / Fax : 0674-2300142 वेब/ Web : <http://www.iopb.res.in>

हमेशा हिंदी में पत्र व्यवहार करके देश का गौरव बढ़ाएं

GENERAL SPECIFICATIONS:-

- | | | |
|--|---|--------------------------------|
| 1. Working Dimension | : | 4'(W) X 2'(D) X 2' (H) |
| 2. Chemical Resistant | : | Outer and Inner coating |
| 3. Storage and Durable Suction | : | 80 to 100 FPM when half closed |
| 4. Vertical lockable and Stable door | | |
| 5. Working top | : | Thick granite |
| 6. Gas nozzle | : | 02 |
| 7. Power sockets for Instruments | | |
| 8. Grills for holding clamps | | |
| 9. Lighting | : | LED or florescent lamp |
| 10. Storage Cabinet at the bottom | : | Must be chemical resistant |
| 11. Ducting and Installation required | | |
| 12. Servicing in case of any emergency quick | | |

Indenter Signature