



भौतिकी संस्थान
भुवनेश्वर
Institute of Physics
Bhubaneswar - 751005, India

(परमाणु ऊर्जा विभाग, भारत सरकार का एक स्वायत्त अनुसंधान संस्थान)
(An autonomous research institution of Dept. of Atomic Energy, Govt. of India)

Revised Quotation Call Notice:

No.IOP/PUR/APEX(SGL)S&M/2022-23/ <u>747</u>	Rate quotation for:
Date: <u>13/07/2023</u>	Supply of Digital Mass flow Controller

To:

M/s

Dear Sir,

In partial modification of this office quotation call notice No. IOP/PUR/APEX/2023-24/699 dated 03.07.2023. 2.Flow measurement accuracy: $\pm 0.06\%$ of reading or $\pm 0.1\%$ of full scale which ever is greater & 3.Flow measurement repeatability: $\pm (0.1\%$ of reading + 0.2% of full scale) may be read as **2.Flow measurement accuracy: $\pm 0.6\%$ of reading or $\pm 0.1\%$ of full scale –whichever is greater & 3.Flow measurement repeatability: $\pm (0.1\%$ of reading + 0.02% of full scale)**

The other terms & conditions remains unchanged.

भवदीय/ Yours faithfully


रजिस्ट्रार/REGISTRAR 13/7/23

डा. घ. - सैनिक स्कूल, भुवनेश्वर - 751 005, भारत

P.O.: SAINIK SCHOOL, BHUBANESWAR - 751 005, INDIA

दूरभाष/ Phone : 0674-2300637, 2301058, 2301823, फैक्स/ Fax : 0674-2300142, वेब/ Web : <http://www.iopb.res.in>

हमेशा हिंदी में पत्र व्यवहार करके देश का गौरव बढ़ाएं



भौतिकी संस्थान
भुवनेश्वर
Institute of Physics

Bhubaneswar - 751005, India

(परमाणु ऊर्जा विभाग, भारत सरकार का एक स्वायत्त अनुसंधान संस्थान)
(An autonomous research institution of Dept. of Atomic Energy, Govt. of India)

QUOTATION CALL NOTICE:

No.IOP/PUR/APEX/2023-24/ <u>699</u> Date: <u>03/07/2023</u>	Rate quotation for: Supply of Digital Mass flow Controller
--	---

To:

Dear Sir,

M/s

Sealed quotation is invited for Supply of Digital Mass flow Controller as per specification, terms and conditions noted below.

Sl. No.	Description/Specification	Quantity (PC)
1	0-200 SCCM std. preloaded gas mass flow meter (along with mounting plate ¼" Swagelok compression end fitting and contactor 1.Measurement: 0.02 to 200SCCM(0.01% to 100% of full scale) 2.Flow measurement accuracy:±0.06% of reading or ±0.1% of full scale – whichever is greater 3.Flow measurement repeatability:±(0.1% of reading +0.2% of full scale) 4.Operating pressure range :11.5 to 160PSIA 5.Pressure reading accuracy:±0.5% of reading or±0.07 PSIA –whichever is greater	01

TERMS & CONDITIONS:

1. Intending supplier should be a registered dealer having GST Registration No. and should submit a copy of the same.
2. Quotation is to be submitted by **post/courier** only in **sealed envelope with superscription "Supply of Digital Mass flow Controller"** and addressed to the Registrar, Institute of physics, P.O: Sainik School, Bhubaneswar-751005.
3. The quotation should be submitted within **21 days** from the date of issue of this letter in the office of the Registrar along with copies of the aforesaid certificates if any failing which the same will be rejected.
4. Rate quoted should remain valid for **90** days from the date of submission of the quotation.
5. Price quoted should be FOR destination at IOP site.
6. Taxes and levies if applicable will be paid extra, which should be specified separately.
7. **No advance payment to be made.**100% payment will be made on delivery, installation of goods in satisfactory conditions and against submission of bill/invoice in duplicate.

भवदीय/ Yours faithfully

(Prof. P.K.Sahu) 3/7/23

रजिस्ट्रार/REGISTRAR

डा. घ. - सैनिक स्कूल, भुवनेश्वर - 751 005, भारत

P.O.: SAINIK SCHOOL, BHUBANESWAR - 751 005, INDIA

दूरभाष / Phone : 0674-2300637, 2301058, 2301823, फैक्स / Fax : 0674-2300142, वेब / Web : <http://www.iopb.res.in>

हमेशा हिंदी में पत्र व्यवहार करके देश का गौरव बढ़ाएं